

Declaration of Conformity

We, 3M Company hereby declare under our sole responsibility that the CE marked products,

3MTM AttestTM Steam Biological Indicator Incubator (56°C), 118 3MTM AttestTM EO Biological Indicator Incubator (37°C), 129

to which this declaration relates are in accordance with the

Electro-Magnetic Compatibility Directive 2014/30/EU, EN 61326-1 (2013); IEC 61326-1 (2012)

and the

Low Voltage Directive 2014/35/EU, IEC 61010-1 (2010) IEC 61010-2-010 (2014)

3M Health Care Business self-declares conformity with Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment and compliance to the requirements of EN 50581:2012.

EU Representative Address

3M Deutschland GmbH Health Care Business Carl-Shurz-Str. 1 41453 Neuss Germany

Signature:

Dianne Gibbs 3M Health Care

Division Regulatory Affairs Manager

Medical Solutions Division

Date: 4 May 2020

Page 1 of 1 Pages

REG-DOC-NMD-05-157083, 3M Attest Biological Indicator Incubators, Version 3, Status: Release, Release Date: 05/01/2020 09:02:06 AM CDT

THIS DOCUMENT WAS PRINTED ON 1-May-2020 9:22 AM